



Schedule

Measurement & Verification Pte Ltd Certificate No. : LA-2015-0594-C

18 Boon Lay Way #03-96 TradeHub 21 Singapore 609966

Date : 15 May 2015

: 1 of 1

Issue No.

Page

Laboratory Location: #03-94

FIELD OF TESTING: Calibration

| | MEASURED QUANTITIES/ RANGE/ INSTRUMENT TO BE CALIBRATED | METHOD | CALIBRATION AND MEASUREMENT CAPABILITY (CMC *) |
|----|---|---|--|
| A. | TEMPERATURE Thermistors 0 °C to 30 °C | In-house procedures : TP 5.4.2-01 (V0) TP 5.4.2-02 (V0) TP 5.4.2-03 (V0) TP 5.4.2-04 (V0) TP 5.4.2-05 (V0) | 0.013 °C |

^{*} CMC is expressed as an expanded uncertainty estimated at a level of confidence of approximately 95%.

Approved Signatories:

Mr Thant Zaw

Mr Lim Chuen Siang

Note:

This laboratory is accredited in accordance with the recognised International Standard ISO/IEC 17025:2005. A laboratory's fulfilment of the requirements of ISO/IEC 17025:2005 means the laboratory meets both the technical competence requirements and **management system requirements** that are necessary for it to consistently deliver technically valid calibration results. The **management system requirements** In ISO/IEC 17025:2005 (Section 4) are written in language relevant to laboratory operations and meet the principles of ISO 9001:2008 **Quality Management Systems** — **Requirements** and are aligned with its pertinent requirements.